

Accreditation



The Deutsche Akkreditierungsstelle attests with this **Accreditation Certificate** that the calibration laboratory

Deutsche WindGuard Wind Tunnel Services GmbH Oldenburger Str. 65, 26316 Varel

meets the requirements according to DIN EN ISO/IEC 17025:2018 for the conformity assessment activities listed in the annex to this certificate. This includes additional existing legal and normative requirements for the calibration laboratory, including those in relevant sectoral schemes, provided they are explicitly confirmed in the annex to this certificate.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and confirm generally with the principles of DIN EN ISO 9001.

This accreditation was issued in accordance with Art. 5 Para. 1 Sentence 2 of Regulation (EC) 765/2008, after an accreditation procedure was carried out in compliance with the minimum requirements of DIN EN ISO/IEC 17011 and on the basis of a review and decision of the appointed accreditation committees.

This accreditation certificate only applies in connection with the notices of 05.10.2023 with accreditation number D-K-15140-01.

It consists of this cover sheet, the reverse side of the cover sheet and the following annex with a total of 2 pages.

Registration number of the accreditation certificate: D-K-15140-01 -00

Berlin, 05.10.2023

Dipl.-Wirtsch.-Ing. (BA) Tim Harnisch Head of Technical Unit Translation issued:

05.10.2023

Dipl.-Wirtsch.-Ing. (BA) Tim Harnisch Head of Technical Unit

The certificate together with the annex reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH (www.dakks.de).

Deutsche Akkreditierungsstelle GmbH

Office Berlin Spittelmarkt 10 10117 Berlin Office Frankfurt am Main Europa-Allee 52 60327 Frankfurt am Main Office Braunschweig Bundesallee 100 38116 Braunschweig

The Deutsche Akkreditierungsstelle GmbH (DAkkS) is the entrusted national accreditation body of the Federal Republic of Germany according to § 8 section 1 AkkStelleG in conjunction with § 1 section 1 AkkStelleGBV. DAkkS is designated as the national accreditation authority by Germany according to Art. 4 Para. 4 of Regulation (EC) 765/2008 and clause 4.7 of DIN EN ISO/IEC 17000.

Pursuant to Art. 11 section 2 of Regulation (EC) 765/2008, the accreditation certificate shall be recognised as equivalent by the national authorities within the scope of this Regulation as well as by the WTO member states that have committed themselves in bilateral or multilateral mutual agreements to recognise the certificates of accreditation bodies that are members of ILAC or IAF as equivalent.

DAkkS is a signatory to the multilateral agreements for mutual recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Co-operation (ILAC).

The up-to-date state of membership can be retrieved from the following websites:

EA: www.european-accreditation.org

ILAC: www.ilac.org IAF: www.iaf.nu



Deutsche Akkreditierungsstelle

Annex to the Accreditation Certificate D-K-15140-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from:

05.10.2023

Date of issue:

05.10.2023

Holder of accreditation certificate:

Deutsche WindGuard Wind Tunnel Services GmbH Oldenburger Str. 65, 26316 Varel

The calibration laboratory meets the requirements of DIN EN ISO/IEC 17025:2018 to carry out the conformity assessment activities listed in this annex. The calibration laboratory meets additional legal and normative requirements, if applicable, including those in relevant sectoral schemes, provided that these are explicitly confirmed below.

The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories and confirm generally with the principles of DIN EN ISO 9001.

Calibration in the fields:

Mechanical quanities Fluid quantities

Velocity of gases

This certificate annex is only valid together with the written accreditation certificate and reflects the status as indicated by the date of issue. The current status of any given scope of accreditation can be found in the directory of accredited bodies maintained by Deutsche Akkreditierungsstelle GmbH at https://www.dakks.de.

Abbreviations used: see last page



Annex to the Accreditation Certificate D-K-15140-01-00

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

	Callbi	11101	i and i	vieasurement Capab	ilities (Civic)	
Measurement quantity / Calibration item		Range	2	Measurement conditions / procedure	Expanded uncertainty of measurement	Remarks
Velocity of gases Absolute value of flow vector Anemometer	0.5 m/s	to	38 m/s	ISO 16622:2002, ISO 17713-1:2007, VA Kalibrierung von Strömungs- geschwindigkeitssensoren (D5831 Version 17)	0.4 %, jedoch nicht kleiner als 0.040 m/s	Wind tunnel: Type Göttingen Nozzle: 1.0 m x 1.0 m, up to 30 m/s additional nozzle: 1.2 m x 1.2 m Anemometer inclination: -31° to 31°
	4 m/s	to	16 m/s	IEC 61400-50-1:2022 section 8, IEC 61400-12-1:2017 Annex F (withdrawn)		
Direction of flow vector Anemometer, wind direction sensor	0°	to	360°	IEC 61400-50-1:2022 Annex A, IEC 61400-12-1:2017 Annex N (withdrawn), ISO 16622:2002, ISO 17713-1:2007, VA Kalibrierung von Strömungsrichtungs- sensoren (D5836 Version 7)	0.8°	Wind tunnel: Type Göttingen Nozzle: 1.0 m x 1.0 m

Abbreviations used:

CMC	Calibration and measu	rement capabilities	(Kalibrier- und N	Messmöglichkeiten)

IEC International Electrotechnical Commission
ISO International Organization for Standardization

VA In-house procedure of Deutsche WindGuard Wind Tunnel Services GmbH

Valid from:

05.10.2023

Date of issue:

05.10.2023